



## Designing Two Synthetic Constructs for Real Time PCR Detection of *Francisella tularensis* and *Ebola* Virus

Mohammadjavad Dehghan Esmat Abadi<sup>1\*</sup>, Hesam Motalebzadeh<sup>2</sup>, Mahmoud Barati<sup>3</sup>,  
and Mohammadali Yaghobi<sup>4</sup>

1. Faculty of Chemistry and Chemical Engineering, Malek Ashtar University of Technology, Tehran, Iran

2. Department of Medical Biotechnology, Faculty of Allied Medical Sciences, Iran University of Medical Sciences, Tehran, Iran

3. Department of Genetics, Faculty of Basic Sciences, Science and Research Branch, Islamic Azad University, Tehran, Iran

4. Faculty of Chemistry and Chemical Engineering, Malek Ashtar University of Technology, Tehran, Iran

### Abstract

**Background:** Generally, timely diagnosis of micro-organisms is very important to prevent many diseases. Many methods can detect micro-organisms like culture-based methods and molecular methods. The molecular methods are usually preferred because they provide fast and reliable results. In some cases, microbial strains are not accessible, and there is no safety to work with them; therefore, synthetic constructs which are designed according to the available sequences in databases can be used as a positive control for detection of them.

**Methods:** In this study, a synthetic construct was designed for molecular detection of *Francisella tularensis* (*F. tularensis*) and the *Ebola* virus by multiplex real-time PCR reaction. For this, sequences were taken from databases and then multiple alignments were done by software. Also, conventional PCR and two models of real-time PCR (SYBR green and TaqMan) were applied. Finally, multiplex real-time PCR was performed.

**Results:** The synthetic construct was designed and used for conventional PCR and multiplex PCR. The results of common PCR showed a single band at 148 bp and 167 bp in 1.5% agarose gel stained by ethidium bromide for *F. tularensis* and *Ebola* virus, respectively. Also, a dual-band at 148 and 167 bp was observed in multiplex PCR. Results of real-time PCR showed a limit of detection about 0.1 pg of plasmid/μl.

**Conclusion:** In conclusion, the designed construct can be used as a positive control for an accurate diagnosis of these micro-organisms without any biological danger for laboratory staff. So, this method is useful for diagnosis of these agents in food, water, and blood samples.

*Avicenna J Med Biotech* 2021; 13(2): 92-97

**Keywords:** *Ebola* virus, Humans, Multiplex polymerase chain reaction, Real-time polymerase chain reaction

### Introduction

*Francisella tularensis* (*F. tularensis*) and *Ebola* virus are highly-infectious elements belonging to family of *Francisellaceae* and *Filoviridae*, respectively<sup>1,2</sup>. *F. tularensis* is a gram-negative intracellular bacterium that causes zoonosis and tularemia or rabbit fever that are highly infectious<sup>2</sup>. These bacteria have two strains including type A strains which are found in America and type B strains which are found in the northern hemisphere<sup>3</sup>. However, these bacteria can survive several days in nature, but they need special conditions to grow up in microbial cultures in the laboratory. Although these bacteria are pathogenic, they can be dan-

gerous for laboratory staff. *Ebola* virus is a non-segmented single-stranded RNA virus which causes hemorrhagic fever<sup>4</sup>. This virus has three species of *Undibugyo*, *Zaire*, and *Sudan*. The first outbreaks of diseases associated with *Ebola* appeared in 1976 in Nzara in Sudan and Yambuku in Congo<sup>5</sup>. According to these features, the existence of a diagnostic method for accurate and safe detection of these pathogens is necessary to prevent infection. There are many diagnostic methods for the detection of bacteria and viruses like culture-based methods, immunological methods, and molecular ones. Cell culture-based methods are

time-consuming and may be dangerous for laboratory staff who work on infectious elements. Immunological methods include Complement Fixation (CF) <sup>6,7</sup>, Enzyme-Linked Immunosorbent Assay (ELISA) <sup>8,9</sup>, and so on. These approaches use antibodies to detect the infectious micro-organisms. These are also time-consuming, because the detection is based on antibody titration and antibodies usually appear nearly 2 weeks after infection. Therefore, molecular methods are more preferred, because of fast and safe diagnosis and their accurate results. In molecular methods such as Polymerase Chain Reaction (PCR) <sup>10-13</sup>, Loop-Mediated isothermal Amplification (LAMP) <sup>14</sup> Nucleic Acid Sequence-Based Amplification (NASBA) <sup>15</sup> and DNA micro-array <sup>16-18</sup>, a specific region in the genomic area can be recognized. These approaches are safe because the genome of micro-organisms is used for experimental works and risk factors affecting laboratory personnel are controlled. Molecular methods like other methods have some problems; for example, LAMP methods for detection are very sensitive and due to their high sensitivity, there is a probability for false positive results or in PCR-based methods, thermocycler is necessary. But the basic problem in these methods is preparing the genome of micro-organisms. In some cases like *F. tularensis* and *Ebola* virus, micro-organism or genome of them is so rare in some countries like our country. Therefore, for solving the problem of diagnosis of infections caused by these pathogens, a new approach is introduced in this study. In this approach, conserved genomic area of micro-organism was selected and then synthesized and cloned to a plasmid vector for amplification <sup>19</sup>. In the current study, a synthetic construct containing a specific genomic area of pathogenic elements (*F. tularensis* and *Ebola*) was designed. This vector was used for a simple PCR reaction for a separate diagnosis of these elements and multiplex PCR reaction for the detection of these elements at the same time. Also, quantitative PCR reaction was applied due to its high specificity and sensitivity. For this study, *fopA* and whole-genome were chosen for the detection of *F. tularensis* and the *Ebola* virus, respectively.

Identification of two pathogens with different clinical symptoms by a positive control structure is the difference between this study and other studies. In fact, the purpose of designing this positive control structure is to obtain the necessary technical knowledge to design a safe positive control for very dangerous pathogens with a high probability of contamination when working with it. This positive control construct also reduces the likelihood of false positives when examining clinical specimens of the patient.

## Materials and Methods

### Synthetic construct designing

In this paper, a specific construct was designed for the fast detection of two micro-organisms containing *F.*

*tularensis* and *Ebola* virus. For this, specific sequences as targets were chosen from NCBI/EMBL. After that, multiple alignments for the selected targets were done by 6Mega BLAST, BioEdit, AliView, and Clustal W software. In the next step, selected sequences were used in SnapGene offline software for designing a plasmid vector containing these sequences. In this step, for each target, two restriction sites were designed at the ends of each target as a simple method for cloning and subcloning the segments. Accordingly, XhoI restriction sites were added to the beginning of *fopA* segment of *F. tularensis* and SacI was added to the end of the *Ebola* segment and both segments were ligated to each other by an EcoRI restriction site. Also, a BamHI restriction site was designed in the middle of segment to prevent false positive results in case of suspicious template. However, the amplicon sizes were different in the construct as well as the actual size in the micro-organisms to prevent vague or false positive results. The designed constructs were shown in figure 1. The base of this pUC57 structure and both pathogen fragments in the relevant plasmid were artificially cloned and lyophilized from the relevant company.

### Designing multiplex primers and hydrolysis probes for multiplex qPCR assay

Target sequences for the reliable detection of *F. tularensis* and *Ebola* virus were selected based on previous literature and sequences available in databases. Selected targets were *fopA* gene (AF097542) for *F. tularensis* and the whole-genome for *Ebola* virus.

Sequences retrieved from NCBI/EMBL were organized and aligned using 6 Mega BLAST software. Oligonucleotides for multiplex qPCR assays and conventional PCR assays were designed using the OligoAnalyzer online software. The design strategy for a multiplex qPCR assay was as follows. First, a hydrolysis probe and primer set were designed for each of the selected target sequences for these agents. Finally, the BLAST N analysis was performed to confirm the specificity of the target organisms.




Figure 1. Designed construct in SnapGene software.

### Construct synthesis and plasmid extraction

After target selection and construct designing, target segments were sent for synthesis of DNA segments and cloning to plasmid vectors then transforming to *Escherichia coli* (*E. coli*) DH5 $\alpha$ . After synthesis of the segments, the bacteria were cultured overnight on LB-broth media containing ampicillin as a selection marker for carrier bacteria. In the next step, plasmid extraction was performed and the extracted constructs were used for PCR amplification.

### Conventional PCR

The common PCR method was performed in a volume of 20  $\mu$ l, containing 10  $\mu$ l of PCR master mix (Amplicon, Denmark), 1  $\mu$ l from each primer (10 pM), 4  $\mu$ l of plasmid as a template (100 pg), and 4  $\mu$ l of sterile water. PCR reaction was performed based on the following program: initial denaturation at 95  $^{\circ}$ C for 5 min, 30 cycles of denaturation at 94  $^{\circ}$ C for 30 s, primer annealing at 62  $^{\circ}$ C for 30 s, extension at 72  $^{\circ}$ C for 30 s, and final extension at 72  $^{\circ}$ C for 5 min. In the end, PCR products were analyzed by 2% agarose gel stained by ethidium bromide. Primer sequences and properties were shown in table 1.

### Multiplex PCR reaction

The multiplex PCR assay was performed via designed primers as follows: 10  $\mu$ l of PCR master mix (Amplicon, Denmark), 1  $\mu$ l from each primer (10 pM), 4  $\mu$ l of plasmid as a template (100 pg), and 4  $\mu$ l of sterile water. In this step, 2 sets of primers were used according to the targets in the plasmid vector.

### Development of real-time PCR

At first, monoplex real-time PCR was done for each segment. In the next step, multiplex real-time PCR was performed by SYBR green. For SYBR green-based qPCR reaction, a mixture with volume of 10  $\mu$ l was prepared as follows: 5  $\mu$ l of SYBR green, 0.5  $\mu$ l from each primer (10 pM), 2  $\mu$ l of plasmid vector as a template (100 pg), and 2  $\mu$ l of sterile water. Finally, multiplex real-time PCR was performed using a designed TaqMan probe. The probes sequences were shown in table 1.

### Sensitivity

The sensitivity of the reaction was tested by making a serial dilution from 100 pg to 0.1 pg of plasmid template for common PCR, multiplex PCR, and real-time PCR assays.

## Results

### Conventional PCR assay

Conventional PCR was performed to confirm the primer and constructs. Also, results showed a specific band at 148 bp and 167 bp for *F. tularensis* and *Ebola* virus, respectively. Also, the sensitivity of the reaction was analyzed using a serial dilution of an extracted plasmid. The results of conventional PCR were shown in figure 2.

### Multiplex PCR assay and sensitivity of the reaction

After designing the primers and construct, a multiplex PCR reaction was performed. According to this, a dual-band at 148 and 167 bp was observed. Also, the sensitivity analysis of the reaction was done and the results were shown in figure 3.

### SYBR green qPCR assay

After the common PCR was done and results confirmed the primers and designed constructs, the SYBR green PCR assay was performed and the results were shown in figure 4.

### TaqMan qPCR assay

In the next step, the segments were used in a TaqMan qPCR reaction with a designed probe. The results were shown in figure 5.

## Discussion

*F. tularensis* and *Ebola* virus are high-infectious elements that cause rabbit fever and *Ebola* hemorrhagic fever. Several methods may detect these microbes. One




Figure 2. A) Analysis of the sensitivity of conventional PCR for *F. tularensis fopA* gene in 1.5% agarose gel stained by ethidium bromide. Lane (M) 100 bp DNA marker, lane (1) negative control, lane (2) 10 ng of DNA, lane (3) 1 ng of DNA, lane (4) 100 pg of DNA, lane (5) 10 pg of DNA, lane (6) 1 pg of DNA, lane (7) 0.1 pg of DNA. B) Analysis of the sensitivity of the PCR reaction for *Ebola* virus in 1.5% agarose gel stained by ethidium bromide. Lane (1) 10 ng of DNA, lane (2) 1 ng of DNA, lane (3) 100 pg of DNA, lane (4) 10 pg of DNA, lane (5) 1 pg of DNA, lane (6) 0.1 pg of DNA, lane (7) negative control and lane (M) 100 bp DNA marker.

Table 1. Primer and probes sequences and properties

Primer and probe	Sequence	Target organism	Amplicon size of construct	Amplicon size of micro-organism
F.T	TGCAGCTAATAATTTTCATTGCTCC	<i>F. tularensis</i>	148 bp	110 bp
R.T	CTACACCTAAGTACCACTGGC			
Probe	TACTTATAGCGCTTTGACTAACAAGGACAATACTTG			
F.E	GATGCCAACGATGCTGTGA	<i>Ebola</i>	167 bp	129 bp
F.T	TGCAAGAGGATGGAGACGAA			
Probe	CAGTGGCTCAAGCTCGTTTTTCAGGTC			




Figure 3. Analysis of the multiplex PCR and sensitivity of the reaction in 1.5% agarose gel stained by ethidium bromide. From left to right, lane (1) 10 ng of DNA, lane (2) 1 ng of DNA, lane (3) 100 pg of DNA, lane (4) 10 pg of DNA, lane (5) 1 pg of DNA, lane (6) 0.1 pg of DNA, lane (7) negative control and lane (M) 100 bp DNA marker.

of the most common diagnostic methods for bacterial detection is culturing, but this approach is not proposed in some cases like *F. tularensis*, because of infectious and fastidious growth conditions. However, there are many serological methods like ELISA, IFA, CFT, and so on for the common detection of both bacterial and viral infections in humans and animals. Berdal *et al* studied the serological methods such as rapid immunochromatography, ELISA, and the common PCR for diagnosis of *F. tularensis* after the outbreak in Norway in 2009<sup>20</sup>. One of the most important problems in these approaches is that they are time-consuming and need special laboratory conditions like the current cases that required BSL3 conditions<sup>21</sup>. Immunological methods work on IgG and IgM antibodies although antibodies may appear 2-3 weeks after the infection. However, these methods are not suitable in some cases like food poisoning, because they are time-consuming. For ex-




Figure 4. A) Analysis of SYBR green real-time PCR of *fopA* gene of *F. tularensis* using Corbett Rotor-Gene 6000. A serial dilution from 100 pg to 0.1 pg of plasmid was examined. B) Analysis of SYBR green real-time PCR of *Ebola* virus using Corbett Rotor-Gene 6000. A serial dilution from 100 pg to 0.1 pg of plasmid was examined.

## Synthetic Constructs for Real Time PCR Detection of *Francisella tularensis* and *Ebola* Virus




Figure 5. A) Analysis of TaqMan real-time PCR of *fopA* gene of *F. tularensis* using Corbett Rotor-Gene 6000. A serial dilution from 100 pg to 0.1 pg of plasmid was examined. B) Analysis of TaqMan real-time PCR of *Ebola* virus using Corbett Rotor-Gene 6000. A serial dilution from 100 pg to 0.1 pg of plasmid was examined.

ample, in these cases, the resulting rate and the reliability of the reaction is very important because the health of water and foods is vital in human life. In past years, the invention of the thermocycler and the appearance of molecular biology approaches established advanced progress in the field of diagnosis and molecular typing of microbes. In this study, a positive control was designed for the rapid and reliable detection of *F. tularensis* and the *Ebola* virus by multiplex real-time PCR reaction. Primers and probes were designed for an accurate diagnosis of the target organisms. In this work, TaqMan probes were used to increase the diagnosis range of different strains. There are so many designed constructs with different goals like detection or identification of some microbes in laboratories or producing subunit vaccines against pathogens. For example, Caasi *et al* designed a positive control for PCR reaction of Barley yellow dwarf virus, Soil-borne wheat mosaic virus, Triticum mosaic virus, and Wheat streak mosaic virus<sup>22</sup>. Also, Pourmahdi *et al* studied the molecular diagnosis of *Yersinia pestis* and *F. tularensis* by multiplex PCR in a synthetic construct and reported the sensitivity of about  $36 \times 10^{-3}$  ng/ $\mu$ l<sup>23</sup>. In another study, Sohni *et al* designed an internal amplification control for the *Bacillus anthracis rpoB* gene, detected by TaqMan real-time polymerase chain reaction and reported a limit of detection about 5 fg/ $\mu$ l<sup>19</sup>. In 1996, Junhui *et al* studied the PCR detection of *F. tularensis* and they compared the sensitivity of the PCR method with the culture-based method to detect bacteria. They chose 3 sets of primers for PCR reaction<sup>24</sup>. In a previous study, Euler *et al* studied real-time PCR detection of *F. tularensis* by a new detection approach. They used a recombinase polymerase for isothermal

real-time PCR amplification and they chose *Tul4* as a target for amplification. Also, they reported a sensitivity of about 102 molecules for their diagnostic method<sup>25</sup>. In the case of detection of the *Ebola* virus, Towner *et al* diagnosed *Ebola* virus by RT-PCR in BSL4 conditions and they reported a limit of detection about 3-38 RNA copies for their method<sup>26</sup>. In another study, Weidmann *et al* studied molecular detection of *Zaire Ebola* virus, *Ebola Sudan* virus and *Marburg* virus (MBGV) by a RT-PCR assay<sup>27</sup>.

### Conclusion

According to the previous study in the field of detection of these micro-organisms and the reported data, our new designed construct and multiplex real-time PCR may be useful for a safe detection of these microbes simultaneously. These tools could be used as diagnosis kits for a rapid and safe detection of these microbes in contaminated or suspicious food and water or dairy products. Also, it could be used in laboratory to prevent the contamination without the need for BSL 3 and 4 laboratory conditions which leads to reduction of laboratory costs.

### References

1. Baden LR, Kanapathipillai R, Campion EW, Morrissey S, Rubin EJ, Drazen JM. Ebola--an ongoing crisis. *N Engl J Med* 2014;371(15):1458-9.
2. Carvalho C, de Carvalho IL, L Zé-Zé L, Nuncio MS, Duarte EL. Tularaemia: a challenging zoonosis. *Comp Immunol Microbiol Infect Dis* 2014;37(2):85-96.
3. Keim P, Johansson A, Wagner DM. Molecular epidemiology, evolution, and ecology of *Francisella*. *Ann N Y Acad Sci* 2007;1105:30-66.

4. Peters C, LeDuc JW. An introduction to Ebola: the virus and the disease. *J Infect Dis* 1999;179 Suppl 1:ix-xvi.
5. Kaushik A, Tiwari S, Jayant RD, Marty A, Nair M. Towards detection and diagnosis of Ebola virus disease at point-of-care. *Biosens Bioelectron* 2016;75:254-72.
6. Owen C, Buker EO, Jellison WL, Lackman DB, Bell JF. Comparative studies of *Francisella tularensis* and *Francisella novicida*. *J Bacteriol* 1964;87(3):676-83.
7. Koskela P, Salminen A. Humoral immunity against *Francisella tularensis* after natural infection. *J Clin Microbiol* 1985;22(6):973-9.
8. Bevanger L, Maeland JA, Naess AI. Agglutinins and antibodies to *Francisella tularensis* outer membrane antigens in the early diagnosis of disease during an outbreak of tularemia. *J Clin Microbiol* 1988;26(3):433-7.
9. Yu JS, Liao HX, Gerdon AE, Huffman B, Scarce RM, McAdams M, et al. Detection of Ebola virus envelope using monoclonal and polyclonal antibodies in ELISA, surface plasmon resonance and a quartz crystal microbalance immunosensor. *J Virol Methods* 2006;137(2):219-28.
10. Long GW, Oprandy JJ, Narayanan RB, Fortier AH, Porter KR, Nacy CA. Detection of *Francisella tularensis* in blood by polymerase chain reaction. *J Clin Microbiol* 1993;31(1):152-4.
11. Sjöstedt A, Eriksson U, Berglund L, Tärnvik A. Detection of *Francisella tularensis* in ulcers of patients with tularemia by PCR. *J Clin Microbiol* 1997;35(5):1045-8.
12. Versage JL, Severin DDM, Chu MC, Petersen JM. Development of a multitarget real-time TaqMan PCR assay for enhanced detection of *Francisella tularensis* in complex specimens. *J Clin Microbiol* 2003;41(12):5492-9.
13. Towner JS, Rollin PE, Bausch DG, Sanchez A, Crary SM, Vincent M, et al. Rapid diagnosis of Ebola hemorrhagic fever by reverse transcription-PCR in an outbreak setting and assessment of patient viral load as a predictor of outcome. *J Virol* 2004;78(8):4330-41.
14. Kurosaki Y, Takada A, Ebihara H, Grolla A, Kamo N, Feldmann H, et al. Rapid and simple detection of Ebola virus by reverse transcription-loop-mediated isothermal amplification. *J Virol Methods* 2007;141(1):78-83.
15. Romano JW, Williams KG, Shurtleff RN, Ginocchio C, Kaplan M. NASBA technology: isothermal RNA amplification in qualitative and quantitative diagnostics. *Immunol Invest* 1997;26(1-2):15-28.
16. Broekhuijsen M, Larsson P, Johansson A, Byström M, Eriksson U, Larsson E, et al. Genome-wide DNA microarray analysis of *Francisella tularensis* strains demonstrates extensive genetic conservation within the species but identifies regions that are unique to the highly virulent *F. tularensis* subsp. *tularensis*. *J Clin Microbiol* 2003;41(7):2924-31.
17. Li Y, Cu YTH, Luo D. Multiplexed detection of pathogen DNA with DNA-based fluorescence nano-barcodes. *Nat Biotechnol* 2005;23(7):885-9.
18. Uda A, Tanabayashi K, Fujita O, Hotta A, Yamamoto Y, Yamada A. Comparison of whole genome amplification methods for detecting pathogenic bacterial genomic DNA using microarray. *Jpn J Infect Dis* 2007;60(6):355-61.
19. Sohni Y, Kanjilal S, Kapur V. Cloning and development of synthetic internal amplification control for *Bacillus anthracis* real-time polymerase chain reaction assays. *Diagn Microbiol Infect Dis* 2008;61(4):471-5.
20. Berdal BP, Mehl R, Haaheim H, Løksha M, Grunow R, Burans J, et al. MeyerField detection of *Francisella tularensis*. *Scand J Infect Dis* 2000;32(3):287-91.
21. Bernard K, Tessier S, Winstanley J, Chang D, Borczyk A. Early recognition of atypical *Francisella tularensis* strains lacking a cysteine requirement. *J Clin Microbiol* 1994;32(2):551-3.
22. Caasi DRJ, Arif M, Payton M, Melcher U, Winder L, Ochoa-Corona FM. A multi-target, non-infectious and clonable artificial positive control for routine PCR-based assays. *J Microbiol Methods* 2013;95(2):229-34.
23. Pourmahdi N, Zeinoddini M, Dehghan Esmatabadi MJ, Sheikhi F. Simple and rapid detection of *Yersinia pestis* and *Francisella tularensis* using multiplex-PCR: Molecular detection of *Yersinia pestis* and *Francisella tularensis*. *Res Mol Med (RMM)*;2019:28-37.
24. Junhui Z, Ruifu Y, Jianchun L, Songle Z, Meiling C, Fengxiang C, et al. Detection of *Francisella tularensis* by the polymerase chain reaction. *J Med Microbiol* 1996;45(6):477-82.
25. Euler M, Wang Y, Otto P, Tomaso H, Escudero R, Anda P, et al. Recombinase polymerase amplification assay for rapid detection of *Francisella tularensis*. *J Clin Microbiol* 2012;50(7):2234-8.
26. Towner JS, Sealy TK, Ksiazek TG, Nichol ST. High-throughput molecular detection of hemorrhagic fever virus threats with applications for outbreak settings. *J Infect Dis* 2007;196(Suppl 2):S205-12.
27. Weidmann M, Mühlberger E, Hufert FT. Rapid detection protocol for filoviruses. *J Clin Virol* 2004;30(1):94-9.