

CRISPR/Cas as a Potential Diagnosis Technique for COVID-19

Coronaviruses are positive single stranded RNA viruses, and are members of Coronaviridae family. Coronaviruses localize in respiratory tract and are usually known as common cold viruses^{1,2}. Seven strains of coronavirus family can infect humans and can cause different signs ranging from cold with major symptoms such as fever and sore throat to upper and lower respiratory tract infections resulting in pneumonia, severe respiratory tract infection and even death. These seven strains include HCoV-229E, HCoV-OC43, SARS-CoV, human coronavirus NL63, human coronavirus HKU1, MERS-CoV and SARS-CoV-2, known as 2019-nCoV or "novel corona virus 2019"³.

At present, "severe acute respiratory syndrome coronavirus 2" or "coronavirus disease 2019" (COVID-19) which is closely related to SARS has become a global health problem. The first-ever COVID-19 case was identified in December 2019 in Wuhan, China; however, since then the virus has spread rapidly across the world and has become a worldwide pandemic and an international concern⁴.

To the best of our knowledge until March 2020, COVID-19 has been reported in 161 countries. COVID-19 is typically transmitted by respiratory droplets during sneezing and coughing⁵. There is no evidence of vertical transmission or transmission during pregnancy^{6,7}. Incubation period of COVID-19 is estimated between 2-14 days and during this time, infected peoples are considered as asymptomatic carriers. Although infection may be asymptomatic, patients typically have fever, cough and shortness of breath. Occasionally disease progresses acutely and causes severe pneumonia, multiple organ failure and death⁸. Patients with underlying medical conditions such as heart and respiratory diseases, asthma, diabetes and immunodeficiency diseases, in addition to elderly age group are high risk and more susceptible to COVID-19 infection⁹. At present, there is no certain treatment or vaccination for prevention of COVID-19 and infected people are either isolated or, in critical conditions, take nonspecific or supportive care⁹. Diagnosis is made based on the symptoms of the disease, chest CT (Computed tomography scan) and qRT-PCR (Quantitative reverse transcription polymerase chain reaction). qRT-PCR technique is the current COVID-19 (SARS-CoV-2) gold standard molecular detection method approved by CDC and World Health Organization (WHO)^{10,11}. Recently, researchers have proposed a coronavirus rapid detection method based on CRISPR/Cas system¹². CRISPR/Cas (Clustered Regularly Interspaced Short Palindromic Repeats) is an adaptive immune system in archaea and bacteria against foreign genetic elements such as phages. Recently, CRISPR/Cas has become a powerful gene editing tool and a promising treatment for genetic diseases and cancers^{13,14}. In this technique,

a programmable protein attaches to the target site by a guide RNA for cleavage of the target sequence. There are several types of Cas proteins that have different properties. Among them, Cas9 protein has received more attention for gene editing whereas, Cas12a and Cas13a have been more efficient in diagnosis of diseases^{15,16}. Cas12a is DNA-specific but Cas13a works with RNA which makes it convenient in detection of SARS-CoV-2. Recently, Zhang *et al* reported specific high-sensitivity enzymatic reporter unlocking (SHERLOCK) technology which is a CRISPR/Cas13 based nucleic acid detection technique for rapid detection of SARS-CoV-2^{17,18}.

They targeted S and ORF1ab protein genes in coronavirus genome. Cas13 identifies and binds to previously determined target sequence which leads to fairly random cleavage of surrounding ssRNA molecules. SHERLOCK technology utilizes a quenched fluorescent ssRNA reporter. The presence of ssRNA coronavirus genome in samples activates Cas13 resulting in the production of quantifiable signals. Amplification of targeted DNA or RNA by Recombinase Polymerase Amplification (RPA) or reverse transcriptase-RPA (RT-RPA) prior to the start of reaction improves the sensitivity of the assay. Subsequently, amplified DNA is converted to RNA by combination of RPA and T7 transcription. Ultimately, detection is made by simultaneous incorporation of the ssRNA reporter (Biotin-RNA-FITC). Viral genome is detected at attomolar concentration using SHERLOCK technology¹⁹. The test can be carried out starting with RNA purified from patient samples, as used for qRT-PCR assays, and can be read out using a dipstick in less than an hour, without requiring elaborate instrumentation¹⁸. As a result, application of CRISPR/Cas13 based diagnosis or SHERLOCK for SARS-CoV-2 detection is much faster than detection by qRT-PCR and has high sensitivity. Consequently, SHERLOCK technology could swiftly replace qRT-PCR technique considering the high demand for rapid diagnosis tests in current global pandemic state of COVID-19.

Avicenna J Med Biotech 2020; 12(3): 201-202

Keywords: Coronavirus, COVID-19, CRISPR/Cas, SARS-CoV-2

References

1. Sexton NR, Smith EC, Blanc H, Vignuzzi M, Peersen OB, Denison MR. Homology-based identification of a mutation in the coronavirus RNA-dependent RNA polymerase that confers resistance to multiple mutagens. *J Virol* 2016;90(16):7415-7428.
2. Battegay M, Kuehl R, Tschudin-Sutter S, Hirsch HH, Widmer AF, Neher RA. 2019-novel Coronavirus (2019-nCoV): estimating the case fatality rate—a word of caution. *Swiss Med Wkly* 2020;150:w20203.

3. Gorbalenya AE. Severe acute respiratory syndrome-related coronavirus—The species and its viruses, a statement of the Coronavirus Study Group. *BioRxiv* 2020.
4. Wang D, Hu B, Hu C, Zhu F, Liu X, Zhang J, et al. Clinical characteristics of 138 hospitalized patients with 2019 novel coronavirus-infected pneumonia in Wuhan, China. *JAMA* 2020;323(11):1061-1069.
5. World Health Organization. Coronavirus disease 2019 (COVID-19): situation report, 45. 2020. https://www.who.int/docs/default-source/coronaviruse/situation-reports/20200305-sitrep-45-covid-19.pdf?sfvrsn=ed2ba78b_4
6. Control CfD, Prevention. Frequently Asked Questions and Answers: Coronavirus Disease 2019 (COVID-19) and Pregnancy. Accessed; 2020.
7. Rasmussen SA, Smulian JC, Lednický JA, Wen TS, Jamieson DJ. Coronavirus Disease 2019 (COVID-19) and Pregnancy: What obstetricians need to know. *Am J Obstet Gynecol* 2020. [Epub ahead of print].
8. Wu Z, McGoogan JM. Characteristics of and important lessons from the coronavirus disease 2019 (COVID-19) outbreak in China: summary of a report of 72 314 cases from the Chinese Center for Disease Control and Prevention. *JAMA* 2020;323(13):1239-1242.
9. Lai CC, Shih TP, Ko WC, Tang HJ, Hsueh PR. Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) and corona virus disease-2019 (COVID-19): the epidemic and the challenges. *Int J Antimicrobial Agents* 2020; 55(3):105924.
10. Jin YH, Cai L, Cheng ZS, Cheng H, Deng T, Fan YP, et al. A rapid advice guideline for the diagnosis and treatment of 2019 novel coronavirus (2019-nCoV) infected pneumonia (standard version). *Mil Med Res* 2020;7(1):4.
11. Shan F, Gao Y, Wang J, Shi W, Shi N, Han M, et al. Lung Infection Quantification of COVID-19 in CT Images with Deep Learning. *arXiv preprint arXiv: 200304655*. 2020.
12. Broughton JP, Deng W, Fasching CL, Singh J, Chiu CY, Chen JS. A protocol for rapid detection of the 2019 novel coronavirus SARS-CoV-2 using CRISPR diagnostics: SARS-CoV-2 DETECTR.
13. Hryhorowicz M, Lipiński D, Zeyland J, Słomski R. CRISPR/Cas9 immune system as a tool for genome engineering. *Arch Immunol Ther Exp (Warsz)* 2017;65(3):233-240.
14. Li Y, Glass Z, Huang M, Chen ZY, Xu Q. Ex vivo cell-based CRISPR/Cas9 genome editing for therapeutic applications. *Biomaterials* 2020;234:119711.
15. Myhrvold C, Freije CA, Gootenberg JS, Abudayyeh OO, Metsky HC, Durbin AF, et al. Field-deployable viral diagnostics using CRISPR-Cas13. *Science* 2018;360(6387):444-448.
16. Li SY, Cheng QX, Wang JM, Li XY, Zhang ZL, Gao S, et al. CRISPR-Cas12a-assisted nucleic acid detection. *Cell Discov* 2018;4(1):20.
17. Metsky HC, Freije CA, Kosoko-Thoroddsen T-SF, Sabeti PC, Myhrvold C. CRISPR-based COVID-19 surveillance using a genomically-comprehensive machine learning approach. *BioRxiv* 2020.
18. Zhang F, Abudayyeh OO, Jonathan SG. A protocol for detection of COVID-19 using CRISPR diagnostics. [https://www.broadinstitute.org/files/publications/special/COVID-19%20detection%20\(updated\).pdf](https://www.broadinstitute.org/files/publications/special/COVID-19%20detection%20(updated).pdf).
19. Kellner MJ, Koob JG, Gootenberg JS, Abudayyeh OO, Zhang F. Author Correction: SHERLOCK: nucleic acid detection with CRISPR nucleases. *Nat Protoc* 2020;15(3):1311.

Mahintaj Dara *, Mahdieh Talebzadeh

Department of Molecular Medicine, Faculty of Advanced Medical Science and Technology, Shiraz University of Medical Sciences, Shiraz, Iran

*** Corresponding author:** Mahintaj Dara, Ph.D., Department of Molecular Medicine, Faculty of Advanced Medical Science and Technology, Shiraz University of Medical Sciences, Shiraz, Iran
E-mail: dara_mp_123@yahoo.com

Received: 14 Apr 2020
Accepted: 27 May 2020